

www.iqproducts.nl

## **PRODUCT INFORMATION SHEET**


Monoclonal antibodies detecting human antigens

Isotype cor IgG2a	EII(	sotype con gG1	trol <sub>R-PE</sub>	REF	IQP-294FR	IVD 50 tests
IVD CE	In Vitro Diagnostic medical device					
	Description					
	IgG2a	Clone	MCG2a	Isotyp	be IgG2a	
	IgG1	Clone	MCG1	Isotyp	be IgG1	
Intended use	Isotype controls are used to detect any non-specific binding, as well as binding mediated by interaction with Fc receptors using flow cytometry.					
Applications	Analysis of non-specific binding in flow cytometry.					
Usage	All these reagents are effectively formulated for direct immunofluorescent staining of human tissue for flow cytometric analysis using 10 $\mu$ l/10 <sup>6</sup> leukocytes for singles and 20 $\mu$ l/10 <sup>6</sup> leukocytes in case of dual and triple combinations. Since applications vary, each investigator should titrate the reagent to obtain optimal results.					

#### **Representative Data**

Staining with clone MCG2a (IgG2a) / MCG1 (IgG1) monoclonal antibodies is illustrated by flow cytometry analysis of normal blood cells. Direct staining was performed using 20  $\mu$ l of the dual and 100  $\mu$ l blood sample.





#### Reproducibility

Monoclonal antibodies from IQ Products were tested by flow cytometry using a 'lyse-wash' method on whole blood from healthy donors. Obtained data support the premise that these reagents are equivalent in their reactivity with peripheral blood lymphocytes. Values are expressed in terms of % of the total lymphocyte count (see table).

Reagent	N	Mean % positive	Product code
IgG2a FITC	10	0,40	IQP-294FR
IgG1 R-PE	10	0,32	IQP-294FR

#### Limitations

- 1. Conjugates with brighter fluorochromes, like PE and APC, will have a greater separation than those with dyes like FITC and CyQ. When populations overlap, the percentage of positive cells using a selected marker can be affected by the choice of fluorescent label.
- 2. Use of monoclonal antibodies in patient treatment can interfere with antigen target recognition by this reagent. This should be taken into account when samples are analyzed from patients treated in this fashion. IQ Products has not characterized the effect of the presence of therapeutic antibodies on the performance of this reagent.
- 3. Reagents can be used in different combinations, therefore laboratories need to become familiar performance characteristics of each antibody in relation with the combined markers in normal and abnormal samples.
- 4. Reagent data performance is based on EDTA-treated blood. Reagent performance can be affected by the use of other anticoagulants.

#### Reagents and materials required but not supplied

- 1. Flow cytometer
- 2. Flow cytometry disposable 12 x 75-mm capped polystyrene test tubes
- 3. Micropipette with disposable tips
- 4. Vortex mixer
- 5. Centrifuge
- 6. IQ Lyse erythrocyte lysing solution (IQP-199)
- 7. IQ Starfiqs fixation and permeabilization solution (IQP-200)
- 8. PBS (phosphate-buffered saline)
- 9. 1% Heparin
- 10. 1% paraformaldehyde solution in PBS (store at 2-8 °C in amber glass for up to 1 week)

### Immunofluorescence staining and lysing protocol

#### Flow cytometry method for use with dual and triple combinations

- 1. Add 100 μl of EDTA-treated blood (i.e. approx. 10<sup>6</sup> leukocytes) to a 5 ml reagent tube. The content of one tube is sufficient to perform one test.
  - For combinations with anti-kappa and/or anti-lambda Ig see application note below.
- 2. Add to each tube 20 µl of labeled monoclonal antibody combination.\*
- 3. Vortex the tube to ensure thorough mixing of antibody and cells.
- 4. Incubate the tube for 15 minutes at room temperature in the dark.
- 5. Add 100 µl of IQ Lyse (IQP-199 ready-to-use) and mix immediately.
- 6. Incubate for 10 minutes at room temperature in the dark.
- 7. Add 2 ml of demineralized water and incubate for 10 minutes in the dark.
- 8. Centrifuge the labeled cell suspension for 2 minutes at 1000 x g.
- 9. Remove the supernatant and resuspend the cells in 200 µl of PBS.\*\*
- 10. Analyze by flow cytometry within four hours (alternatively, the cells may be fixed by 0.05% of formaline in buffered saline for analysis the next day. Some antigens are readily destroyed upon fixation and this should be taken into account when using this alternative).

\* Appropriate mouse Ig isotype control samples should always be included in any labeling study \*\* PBS: Phosphate Buffered Saline, pH 7.2



Add 2 ml of PBS containing 0.001% (v/v) Heparin (**prewarmed to 37 °C**) to the cell suspension Vortex, centrifuge (2 min at 300x g) and discard the supernatant Repeat this step twice Resuspend the pelleted blood cells in 100 μl PBS containing 0.001% (v/v) Heparin

CE

# ⚠ 🙈 🔏 🛣 🗳

## Handling and Storage

Antibodies are supplied either as 100 tests per vial (1 ml) for singles or 50 tests per vial (1 ml) for dual and triple combinations. They are supplied in 0.01 M sodium phosphate, 0.15 M NaCl; pH 7.3, 0.2% BSA, 0.09% sodiumazide (NaN<sub>3</sub>). Store the vials at 2-8 °C. Monoclonal antibodies should be protected from prolonged exposure to light. Reagents are stable for the period shown on the vial label when stored properly.

**Warranty** Products sold hereunder are warranted only to conform to the quantity and contents stated on the label at the time of delivery to the customer. There are no warranties, expressed or implied, which extend beyond the description on the label of the product. IQ Products is not liable for property damage, personal injury, or economic loss caused by the product.

#### Characterization

- To ensure consistently high-quality reagents, each batch of monoclonal antibody is tested for conformance with characteristics of a standard reagent. Representative flow cytometric data is included in this data sheet.
- **Warning** All products contain sodiumazide. This chemical is poisonous and hazardous. Handling should be done by trained staff only.

#### Explanation of used symbols

Explanatio	n or used symbols
<b>II</b>	Consult instructions for use
REF	Catalogue number
$\overline{\mathbb{V}}$	Sufficient for
IVD	In Vitro Diagnostic medical device
$\overline{\mathbb{A}}$	Caution, consult accompanying document
*	Keep away from (sun)light
æ	Biological risks
1	Temperature limitation (°C)
RUO	For Research Use Only
LOT	Batch code
2	Use by yyyy-mm-dd
	Manufacturer
EC REP	Authorized Representative in the European Community
CE	Conformité Européenne (European Conformity)
	Dright tillorescence

		Label - tandem	Ex -max (nm)	Em -max (nm)
Р	PURE	purified material	-	-
F	FITC	FITC	488	519
R	R-PE	PE	488, 532	578
С	CyQ	PE-Cy5.18	488, 532	667
Α	APC		595, 633, 635, 647	660
PC	PerCP		488, 532	678
PCC	PerCP-Cy5.5		488, 532	695

IQ Products BV

Rozenburglaan 13a 9727 DL Groningen, The Netherlands

- +31 (0)50 57 57 000
- ∃ +31 (0)50 57 57 002
- Technical <u>marketing@iqproducts.nl</u>
- Orders <u>orders@iqproducts.nl</u>
- <u>www.iqproducts.nl</u>