

## DECLARATION OF CONFORMITY

<b>Date of issue</b>	Not applicable	<b>Valid until</b>	Not applicable
<b>CE Assessment certificate</b>	This product has been assessed on conformity according to the essential requirements as listed in Annex I of the European Directive 98/79/EC on in vitro diagnostic medical devices (IVDD). Certification by a Notified Body is not required.		
<b>Certificate no.</b>	Not applicable		

<b>Product code</b>	QQF-100	<b>Product name</b>	FMH QuikQuant™
<b>EDMA code</b>	130190.90	<b>EDMA term</b>	Other Hematology Reagents
<b>EMDN code Level 4</b>	W010301	<b>EMDN term</b>	Haematology Reagents
<b>GMDN code</b>	55975	<b>GMDN term</b>	Foeto-maternal haemorrhage/Kleihauer test IVD, kit, fluorescent activated cell sorting/flow cytometry (FACS/Flow)
<b>Basic UDI - DI</b>	87179530223QQF-100N8		
<b>Intended use</b>	The FMH QuikQuant™ is intended for the discrimination and quantitative detection of human fetal red blood cells (fRBC) in maternal blood. This method used for diagnosis of Fetomaternal Hemorrhage (FMH) is applied to peripheral blood samples of pregnant women with abdominal trauma and/or suspected Rhesus D (RhD) incompatibility. The FMH QuikQuant™ is based on a sensitive and accurate, non-automated flow cytometric method, which offers a fluorescent detection of the intracellular antigen, fetal hemoglobin (HbF). HbF is detected in red blood cells obtained from EDTA anti-coagulated or heparin-treated human peripheral whole blood. The FMH QuikQuant™ is intended for use in hospital clinical and reference laboratories by trained medical technologists or similar individuals having experience in test methods for FMH and flow cytometry.		

<b>Manufacturer name and Address</b>	IQ Products BV, Rozenburglaan 13a, 9727 DL Groningen, The Netherlands
<b>Single Registration number (SRN)</b>	NL-MF-000010812
<b>Competent Authority</b>	Dutch Health Care Inspectorate CIBG, registration no. NL-CA002-39072
<b>Harmonized Standards Referenced</b>	<ul style="list-style-type: none"> <li>• EN ISO 13485:2016      • EN ISO 18113-1:2011</li> <li>• EN 13641:2002        • EN ISO 18113-2:2011</li> <li>• EN 13612:2002        • EN ISO 23640:2015</li> <li>• EN 13975:2003        • EN ISO 15223-1:2021</li> <li>• EN ISO 14971:2012</li> </ul>
<b>Other regulations/standards by which the product is regulated</b>	Regulation (EC) No. 1272/2008; Regulation (EU) No. 453/2010; Regulation (EU) No. 2015/830.

This declaration of conformity is issued under the sole responsibility of IQ Products BV.

- (1) The product meet the provisions of the Directive 98/79/EC on in vitro diagnostic medical devices which apply to the products;
- (2) The FMH QuikQuant™ is placed on the market for the first time in 2012;
- (3) The product concerned is produced at the manufacturing facility of IQ Products B.V. in conformance with the design control procedure requirements as specified in Directive 98/79/EC and the records are available for review;
- (4) The product comply with all essential requirements of the Directive 98/79/EC on in vitro diagnostic medical devices;
- (5) IQ Products BV has a Quality System in place based on ISO 13485:2016. The Quality System has been certified and the certificate has been issued by Notified Body DEKRA Certification B.V. (Identification number 0344) on February 01, 2022 and expires on February 01, 2025;

We hereby certify that as of the date of this declaration, the product described above conform with the provisions of the European Directive 98/79/EC on in vitro diagnostic medical devices (IVDD). All supporting documentation is retained at IQ Products BV.

This declaration of conformity is issued under the sole responsibility of IQ Products BV

Date: \_\_\_\_\_

Signature: \_\_\_\_\_

Place: Groningen, the Netherlands

Name: Christine M. Kebbedies

Function: Manager QA & RA

Company: IQ Products BV