



CMV Brite™ Turbo Kit

*Rapid CMV pp65 Antigenemia for the
detection of active CMV infection*



Package Insert



110 Tests

[REF]⁹ VIR-CMV 110 [REF]¹⁰ VIR-CMV 110 BDC

[IVD] **CE0344 In Vitro Diagnostic medical device**

[IVD] 510(k) #991650

[UDI-DI] 87179530223VIR-CMVZC

Barcode GS1 (GTIN): 8717953022332

INTERNATIONAL PACKAGE INSERT ENGLISH - PORTUGUES

English **3**

Portugues **6**

Other languages of this Product's Instruction for Use (IFU) / Product Insert can be found at the World Wide Web at www.iqproducts.nl. Copies may also be requested by sending an email request to marketing@iqproducts.com or by contacting your local product distributor.

CMV Brite™ Turbo Kit

Rapid CMV pp65 Antigenemia for the detection of active CMV infection

REF¹⁰ VIR-CMV 110 ▽ 110 Tests **IVD** **CE0344**

REF¹⁰ VIR-CMV 110 BDC ▽ 110 Tests **IVD** 510(k) #991650
complete test package

Intended use

The CMV Brite™ Turbo Kit is intended for the rapid qualitative detection of Cytomegalovirus (CMV) lower matrix protein pp65 by indirect immunofluorescence using microscopy in isolated peripheral blood leukocytes obtained from ethylenediaminetetraacetic acid (EDTA) or heparin anti-coagulated human peripheral blood. The detection of CMV pp65 in human peripheral blood cells aids in the diagnosis of acute or reactivated CMV infection. This product is not FDA cleared (approved) for use in testing (i.e. screening) of blood or plasma donors.

Summary and Explanation

Infections with human cytomegalovirus (CMV), a β-herpes virus, are widespread throughout the world (rate of prevalence 40-100%). While an infection with CMV proceeds asymptotically in the majority of immunocompetent persons, it can lead to serious complications in persons whose immune system has been weakened or is not yet fully developed (hepatitis, retinitis, pneumonia etc.). CMV is the most frequent pathogen of congenital infections. Approximately 10% of children congenitally infected with CMV show symptoms at birth (icterus, hepatosplenomegaly, petechial bleeding and chorioretinitis). Further groups of patients for whom an acute CMV infection represents a serious threat are recipients of organ and bone-marrow transplants and AIDS patients. The CMV antigenemia assay has been developed using a cocktail of two monoclonal antibodies (C10/C11) directed against pp65 [2]. The CMV antigenemia assay is valuable in the diagnosis of active CMV infection. Shell vial cultures provide a result within 1 to 2 days, but are not sensitive for detection of CMV in blood specimens. In contrast, detection of CMV antigen in peripheral blood polymorphonuclear (PMN) cells (CMV antigenemia) is both sensitive and rapid [1, 2]. This technique uses monoclonal antibodies to detect the CMV lower matrix phosphoprotein (pp65), an early antigen in virus replication, that is abundantly present in antigen-positive PMNs [2-5]. The CMV Brite Turbo assay can be completed within 2 hours of sample collection.

Principle of the test - immunofluorescence test -

The CMV Brite™ Turbo method consists of:

- a. direct lysis of peripheral blood erythrocytes
- b. preparation of cytocentrifuge slides
- c. fixation and permeabilization
- d. indirect immunofluorescence staining using monoclonal antibodies directed

against CMV pp65 protein.

- e. reading and evaluation of results

The first step in the CMV Brite™ Turbo method involves direct lysis of the peripheral blood erythrocytes [6]. Following lysis the leukocytes are cytocentrifuged onto a slide, fixed and permeabilized to allow subsequent detection of pp65 antigen. The presence of pp65 is detected by the C10/C11 antibody cocktail and visualized by means of a specific secondary FITC-labeled antibody. CMV antigen-positive leukocytes exhibit homogeneous yellow-green polylobate nuclear staining when observed using a fluorescence microscope. The number of CMV antigen-positive cells is counted per duplicate stain.

Kit contains

LYS	200 ml	Reagent A (A) , Erythrocyte lysing solution (Ammonium chloride solution, sodium azide < 0,1%) Concentrate: dilute 1:10 with demineralized water
FIX	290 ml	Reagent B (B) , Fixative solution (formaldehyde in PBS, sodium azide < 0,1%) Concentrate: dilute 1:5 with PBS
PER	290 ml	Reagent C (C) , Permeabilization solution (Igepal Ca-630, newborn calf serum in PBS, sodium azide < 0,1%), Concentrate: dilute 1: 5 with PBS
MAB	4 ml	Reagent D (D) , Monoclonal antibody (mouse), Mix of C10/C11 (IgG/IgG) against lower matrix protein pp65. sodium azide < 0,1% (Ready to use)
CONJ	4 ml	Reagent E (E) , FITC-conjugated sheep anti-mouse-immunoglobulins with Evans Blue, sodium azide < 0,1%. (Ready to use)
CONTR	5 x 1	Control Slide , CMV antigenemia control microscope slides. Control slide in a sealed pouch with desiccant. (Ready to use)

Laboratory material required and not included in the Kit

Laboratory centrifuge; 50 ml conical bottomed centrifuge tubes; Sterile pipettes, micropipette and tips; phosphate buffered saline (PBS), pH 7.4, Ca, Mg free; Hemocytometer or automated cell counter; Cytocentrifuge slides; Cytocentrifuge (such as Shandon Southern Products, Ltd., model Cytospin 3); Coplin jars or histology staining jars; Humid chamber; 37 °C incubator; Fluorescence microscope capable of 250x to 1000x magnification; Mounting medium (non-fluorescent, such as Citifluor, Glycerol PBS solution, UKC Chem Lab; or Immunoconcept Mounting Medium, Catalogue # 111); Fume Hood; Micro cover glass; Stop watch/Timer.

Warning and precautions

Do not incorporate reagents. Avoid contact with eyes and skin. All samples and materials used for the test must be treated as being potentially infectious and appropriate safety precautions taken. In the preparation of the CMV Brite™ Turbo Control Slides, leukocytes have been used obtained from a healthy human blood donor. Each donor sample has been tested and found non-reactive for the presence of antibodies to HIV-1 , HIV-2, HCV and CMV as well as for HBsAg. Do not pipette with mouth. According to good laboratory practice wear gloves, laboratory coat and safety glasses. Liquids and non-combustible materials should be decontaminated with sodium hypochlorite (final concentration: 3%, activity time at least 30 minutes). Liquid waste which contains acids must be neutralized before disposal. All materials that are to be reused must be autoclaved for 1 hour at 121 °C.

LYS contains Ammonium chloride. H302 Harmful if swallowed; H319 Causes serious eye irritation. P305 + P351 + P338 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. **FIX** contains less than 9.3% formaldehyde. Formaldehyde is a highly toxic allergenic and potentially carcinogenic reagent, which should be handled in accordance with good laboratory practices using appropriate precautions. Avoid skin or eye contact. H301 + H311 + H331 Toxic if swallowed, in contact with skin or if inhaled; H314 Causes severe skin burns and eye damage; H317 May cause an allergic skin reaction; H335 May cause respiratory irritation; H351 Suspected of causing cancer; H370 Causes damage to organs. P260 Do not breathe dust/fume/gas/mist/vapours/spray; P280 Wear protective gloves/ protective clothing/ eye protection/ Face protection; P301 + P310 IF SWALLOWED: Immediately call a POISON CENTER or doctor/physician; P305 + P351 + P338 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing; P310 Immediately call a POISON CENTER or doctor/ physician. **CONJ** contains Evans Blue. Avoid contact with eyes, skin, clothing. Keep container tightly closed. Wash thoroughly after contact with **FIX** and **CONJ** and consult a doctor. The test must be performed by well-trained and authorized laboratory technicians. Testing is performed under aseptic and microbiologically controlled conditions. Please contact the manufacturer if the original test kit is damaged.

Storage

Upon receipt, store reagents at 2-8 °C. Avoid direct sunlight. Reagents stored according to stated storage instructions are stable until the expiration date indicated on the label. For repeatedly testing store the reagents immediately after usage at 2-8 °C.

Processing of the blood sample

Collect between 3 to 5 ml venous blood into an EDTA-treated tube, using aseptic venipuncture. Send the sample to the laboratory without delay. The blood sample should be kept at room temperature (20-25 °C) until processing. Processing should be performed within 6 to 8 hours of sample collection since it has been shown that a decrease of antigen-positive cells (collected in heparin) can be found after storage [7-8, 11]. Therefore, immediate preparation of the slides should always be attempted. In patients with severe neutropenia (absolute neutrophil count less than 200/mm³) at least 10 ml of blood may be required. If samples are to be transported, they must be packed in accordance with legal requirements for the transportation of infectious materials.

Test Procedure

The protocol has to be followed strictly. Unless stated otherwise, reagents (including PBS) should be at room temperature when used in the test procedure. Room temperature is defined as 20-25 °C. Allow control slides to equilibrate to room temperature prior to opening.

I. Preparation of leukocyte suspension

- Dilute Reagent A 1:10 in demineralized (distilled) water and allow to cool to 2-8 °C.
- Mix 2 ml blood with 30 ml of cold (+4 °C) diluted erythrocyte lysing solution in a 50 ml conical bottomed tube and incubate for 5 minutes at 2-8 °C.
- Centrifuge for 2 minutes at 1000xg (2500 rpm). Discard the supernatant.
- Repeat the cold lysing step if lysing of erythrocytes is not sufficient after the first time.
- Resuspend the cell pellet in 30 ml PBS.
- Centrifuge for 2 minutes at 1000xg (2500 rpm). Discard the supernatant
- Resuspend the pellet in 1 ml PBS. (In patients with severe neutropenia re-suspension in as low as 0.2 ml may be required).

II. Cell counting

- Count cells using a Hemocytometer or automated cell counter.
- Adjust concentration to 2.0×10^6 cells/ml by diluting in PBS.

III. Preparation of cytocentrifuge slides

- Centrifuge 100 µl of the 2.0×10^6 cells/ml suspension at approx. 600 rpm (54xg) for 4 minutes onto glass slides by cytocentrifuge.
- Prepare at least 3 slides per patient specimen (2 for testing, plus an additional slide for backup).
- Let slides dry for approximately 5 minutes.
- Circle cell area on the slide using a laboratory marker pen.
- Slides can be kept at room temperature overnight before fixation.

IV. Fixation and permeabilization

- Dilute fixative (reagent B) 1: 5 in PBS in a fume hood prior to use.
- Dilute permeabilization solution (reagent C) 1: 5 in PBS in a fume hood prior to use. (Do not reuse.)
- Immerse 2 slides in diluted reagent B for 5 minutes at room temperature in the fume hood.
- Dip slides 3 times in PBS (washing solution) and leave in the washing solution for 3 minutes.
- Immerse slides in diluted reagent C for 1 minute at room temperature.
- Dip slides 3 times in washing solution and place in fresh washing solution for 5 minutes (or any time up to 60 minutes). If staining with monoclonal antibody is to follow directly then proceed to step V.

To store slides:

If slides are to be stored then rinse in demineralized (distilled) water for 15 seconds. Let slides dry under the ventilator for approximately 20 minutes. Once dry, slides should be packed in aluminum foil and stored at 2-8 °C for 24 hours, or frozen at -80 °C.

V. Immunofluorescence staining

- From this point on, do not allow the cell preparations to dry out during the remainder of the staining procedure.
- CMV Brite™ Turbo Control slide. Rehydrate control slide in PBS for 1 or 2 minutes.
- Remove one slide at a time from the washing solution, carefully dry the area surrounding the cell spot.
- Apply 35 µl of C10/C11 MoAb solution (reagent D) incubate for 20 minutes at 37 °C in a humid chamber.
- Dip slides 3 times in washing solution (PBS) and put in fresh washing solution for 3 minutes.
- Remove one slide at a time from the washing solution, carefully dry the area surrounding the cell spot.
- Apply 35 µl conjugate (reagent E), incubate for 20 minutes at 37 °C in a humid chamber.
- Wash twice in fresh PBS and carefully rinse with tap water (3 times) and mount with mounting medium and a micro cover glass.

VI. Reading

Perform reading as soon as possible. Slides may be stored for up to 8 hours at 2-8 °C covered tightly in order to minimize fading. Perform microscopic evaluation using an immunofluorescence microscope at 400x magnification. A higher magnification of 1000x may be used to increase resolution. (See note on recommendations for microscope and fluorescence cube under "Laboratory material and equipment".) Scan the whole surface of the spot. Two spots should be scanned per patient. Positive cells show homogenous yellow-green polylobate nuclear staining. Negative cells show no yellow-green nuclear staining. Equivocal readings due to artifacts may occur rarely (less than 5%).

The most common artifact is due to eosinophils. They are recognized by a nonspecific cytoplasmatic staining, with the nucleus appearing as black holes, often spectacle-shaped. The cytoplasm may appear dull and more yellow. If all cells appear greenish, this represents an artifact, which may be associated with some types of patients and is very rare.

A greenish staining of cells at the periphery of the spot only, is not considered positive. This may occur if the spot has started to dry out during incubation with either monoclonal antibody solution or conjugate solution.

Be sure to incubate in a humid chamber and check that the whole cell spot is covered with the reagent.

If slides are not interpretable due to equivocal readings of artifacts, stain backup slide(s) or if unresolved obtain and test an additional specimen.

Quality Control

The positive (fixed CMV lower matrix protein pp65 antigen-positive cells mixed with antigen-negative leukocytes) and negative (fixed CMV antigen-negative leukocytes) control slides are provided with the kit. The slides are only used to check the staining procedure and do not influence the diagnostic value of the kit. The positive control must demonstrate appropriate staining before patient specimens can be evaluated. The positive control should exhibit homogeneous yellow-green staining of positive cells with round morphology (nucleus is not visible). The negative control should show no yellow-green staining. Despite the fact that the quality procedures are followed strictly, single positive cells could be observed in the negative control. In such cases the staining run should **not** be considered invalid. The single positive cells on the negative control do not mean that the slide with patient material cannot be interpreted.

Choice of microscope

Selecting the correct filter cube configuration, which matches the fluorochrome in use, is important. This will vary for each microscope manufacturer. Select the correct combination for FITC use, such as the Olympus BX-40 or BX-60 microscope with: Fluorescence cube (U-MNB), Excitation wavelength (470-490 nm), Emission wavelength (> 515 nm).

Reading and interpretation of result

Results of evaluation of patient specimen slides are qualitative.

Positive result: one or more CMV antigen-positive cells present per duplicate stain. **Negative result:** no CMV antigen-positive cells present per duplicate stain. A minimum of approximately 50,000 specimen cells should be present in order to determine that a result is negative. If equivocal readings due to artifacts occur on both duplicate stains, the results should not be interpreted. Stain back-up slides or obtain and test an additional patient specimen.

Limitations of the procedure

Detection of CMV pp65 lower matrix, early structural protein should be performed by laboratories experienced in immunocytochemical techniques.

Leukocyte preparation should be performed by personal experienced in aseptic techniques. The efficacy of the CMV Brite™ Turbo Kit with samples other than human blood leukocytes has not been established. Test performance characteristics have been established using the CMV Brite™ Kit (EDTA and heparin specimens) and validated in internal studies using the CMV Brite™ Turbo Kit (EDTA specimens only). The CMV Brite™ Turbo Kit is intended for use only with immunofluorescent microscopy and not for use with flow cytometry. The type and condition of the instrumentation used will influence the visual appearance of the image obtained. The reaction may vary due to the type of microscope employed, the light source, age of the bulb, filter assembly and filter thickness, differences in sensitivity of the antigen substrate, or the assay procedure used. Each laboratory should establish its own criteria for the reading of patient specimens using appropriate controls. The detection and confirmation of CMV pp65 lower matrix, early structural protein in peripheral blood leukocytes is not diagnostic of symptomatic illness, since CMV pp65 may be present for a significant period following acute infection and patients with CMV antigenemia (especially low levels) may have asymptotic illness. There are many published reports of patients who are antigenemia-negative and viremia-positive and some of these patients may have CMV disease. Thus, a negative antigenemia test does not absolutely exclude CMV infection or disease. The CMV Brite™ Turbo Kit is not intended for antiviral drug monitoring. Test performance characteristics have not been established using neonatal specimens. A decrease can be noted of antigen-positive cells (collected in heparin) per slide after storage [7-8, 11]. Therefore, immediate preparation of the slides should always be attempted. The maximal specimen storage time at room temperature has not been determined. In patients with severe neutropenia (absolute neutrophil count less than 200/mm³) at least 10 ml of blood may be required. Since the monoclonal antibodies have been prepared using a prototype strain, they may not detect all antigenic or new strains of CMV. For example, monoclonal antibodies may fail to detect strains of CMV which have undergone minor amino acid changes in the target epitope region. Test performance characteristics have not been established using specimens collected in heparin.

Performance characteristics

The CMV Brite™ Kit was compared to CMV virus detection in conventional culture (CC) and shell vial culture (SV) using human peripheral blood leukocytes. 300 clinical samples were evaluated in comparison to conventional culture (CC) and shell vial culture (SV) [9]. A total of 300 venous blood samples were collected from transplant recipients (159 allogeneic bone marrow and 47 solid organ), 85 HIV positive/AIDS patients, and 9 immunocompetent patients. 92 samples (30.7%) were determined to be CMV positive by the CMV Brite™ Kit. In random blood donor populations, the number of specimens found repeatedly reactive for CMV antigenemia by the CMV Brite™ Kit has typically been 0%. Note: The performance characteristics have been established using the CMV Brite™ Kit and validated in internal studies using the CMV Brite™ Kit Turbo Kit.

Comparison of CMV Brite™ Kit with CC and SV

CC/SV	CMV Brite™ Kit		total
	+	-	
+	31	3	34
-	61	205	266
total	92	208	300

Sensitivity: 31/34 = 91.2% (95% CI = 76.3 to 98.1%)

Specificity: 205/266 = 77.1% (95% CI = 70.2 to 81.1%)

The performance of the CMV Brite™ Turbo Kit was validated in a study of 183 patient samples. The patient samples included 173 organ transplant patients, 9 immunocompetent patients and 1 HIV positive patient sample. Each patient was tested in parallel with the CMV Brite™ and the CMV Brite™ Turbo Kits. The results of this validation study are provided in the table below.

CMV Brite™ Turbo Kit	CMV Brite™ Kit		
	+	-	Total numbers
+	43	7	50
-	6	127	133
Total numbers	49	134	183

Cross reactivity

For the analysis of cross reactivity clinical isolates of the following viruses were tested: Herpes virus type 1 and type 2, Varicella-zoster virus, Adenovirus 2, 4 and 5, Parainfluenza virus 1, 2 and 3, Respiratory syncytial virus, Poliovirus 3 (wild type 3), Echovirus 11, Echovirus 30, Epstein-Barr virus (laboratory strain P3HR1) , Human herpes virus type 6 (laboratory strain GS) , Human immunodeficiency virus (type 1 commercially available IF slides). No cross reactivity was observed except for a weak positive reaction with six of the seven HSV-1 isolates in the shell vial assay. The weak staining observed with the six HSV-1 isolates was cytoplasmic (i.e., seen only outside the nucleus, in a limited number of small foci), similar to the IF pattern usually found with HSV-1 monoclonal antibodies. The weak staining observed may have been due to Fc receptors being expressed by the HSV-1 infected cells in the shell vial assay. Subsequent analysis concluded that there is no evidence of cross reactivity between HSV and the CMV antigenemia assay using the C10/C11 monoclonal antibody cocktail.

Regulatory Status

At this time, the CMV Brite™ Turbo Kit is registered as "in vitro diagnostic medical device" in Brasil, South Korea, the United States of America and in the countries that belong to the European Community. In all other countries it should be labeled "for research use only".

Literature

1. Van der Bij, W., Schirm, J., Torensma, R., Van Son, W.J., Tegzess, A.M., The, T.H. (1988) Comparison between viremia and antigenemia for detection of cytomegalovirus in blood. *J. Clin. Microbiology* **26**, 2531 - 2535.
2. Grefte, J.M.M., Van der Gun, B.T.F., Schmolke, S., Van der Giessen, M., Van Son, W.J. and The, T.H. (1992). The lower matrix protein pp65 is the principal viral antigen present in peripheral blood leukocytes during an active cytomegalovirus infection. *J.Gen.Viro.* **73**, 2923-2932.
3. Gerna, G., Revello, M.G., Percivalle, E., Zavattoni, M., Parea, M., Battaglia, M. (1990). Quantification of human cytomegalovirus viremia by using monoclonal antibodies to different viral proteins. *J. Clin. Microbiology* **28**, 2681 - 2688.
4. Revello, M.G., Percivalle, E., Di Matteo, A., Morini, F., Gerna, G. (1992). Nuclear expression of the lower matrix protein of human cytomegalovirus in peripheral blood leukocytes of immunocompromised viremic patients. *J. Gen Virol* **73**, 437 - 442.
5. Gerna, G., Revello, M.G., Percivalle, E., Morini, F. (1992). Comparison of different immunostaining techniques and monoclonal antibodies to the lower matrix phosphoprotein (pp65) for optimal quantitation of human cytomegalovirus antigenemia. *J. Clin. Microbiology* **30**, 1232 - 1237.
6. Ho, S.K.N., Lo, C-Y., Cheng,I.K.P. and Chan,T-M., (1998) Rapid cytomegalovirus pp65 antigenemia assay by direct Erythrocyte Lysis and Immunofluorescence Staining. *J. Clin.Microbiology* **36**, 638-640.
7. Boeckh, M., Woogerd, P.M., Stevens-Ayers, T., Ray, C.G., and Bowden, R.A. (1994). Factors influencing detection of quantitative Cytomegalovirus antigenemia. *J.Clin.Microbiology* **32**, 832-834.
8. Landry, M.L., Ferguson D., Cohen, S., Huber, K., and Wetherill, P. (1995). Effect of delayed specimen processing on Cytomegalovirus antigenemia test results. *J.Clin.Microbiology* **33**, 257-259.
9. Landry, M.L., Ferguson, D., Stevens-Ayers, T., De Jonge, M.W.A., and Boeckh, M. (1996). Evaluation of CMV Brite™ Kit for detection of Cytomegalovirus pp65 antigenemia in peripheral blood leukocytes by immunofluorescence. *J.Clin.Microbiology* **34**, 1337-1339.
10. NEN EN ISO 15223-1 Medical devices - Symbols to be used with medical device labels, labeling and information to be supplied – Part 1: General requirements.
11. Niubò, J., Pérez, J.L., Carvajal, A., Ardanuy, C. and Martín, R. (1994). Effect of delayed processing of blood samples on performance of cytomegalovirus antigenemia assay. *J.Clin.Microbiology* **32**, 1119-1120.

Explanation of used symbols

	Consult instructions for use
	Catalogue number
	Sufficient for
	In Vitro Diagnostic medical device
	Caution, consult accompanying document
	Keep away from (sun)light
	Biological risks
	Temperature limitation (°C)
	For Research Use Only
	Batch code
	Use by yyyy-mm-dd
	Manufacturer
	Authorized Representative in the European Community
	Conformité Européenne (European Conformity)

Contact information**IQ Products BV**

www.iqproducts.nl

Rozenburglaan 13a
9727 DL Groningen
The Netherlands
T +31 (0)50 5757000
F +31 (0)50 5757002
marketing@iqproducts.nl

CMV Brite™ Turbo Kit

Antigenemia rápida para pp65 CMV para detecção de infecção activa por CMV

REF¹⁰ VIR-CMV 110 ▼110 Testes **IVD** CE0344

REF¹⁰ VIR-CMV 110 BDC ▼ 110 Tests **IVD** 510(k) #991650

Finalidade de utilização

O CMV Brite™ Turbo Kit destina-se à rápida detecção qualitativa da proteína pp65 da matriz inferior do Citomegalovírus humano (CMV) por imunofluorescência indireta através de microscopia em leucócitos isolados de sangue periférico obtidos de sangue periférico humano anticoagulado com ácido etilenodiaminotetracético (EDTA) ou heparina. A detecção de CMV pp65 em células sanguíneas periféricas humanas auxilia para o diagnóstico da infecção aguda ou reactivada por CMV. Este produto não é aclarado pelo FDA (aprovada) para utilização em ensaios (por exemplo, triagem) dos dadores de sangue ou plasma.

Resumo e explicação

As infecções por CMV (citomegalovírus humano) um vírus β-herpes, estão disseminadas por todo o mundo (taxa de prevalência de 40-100%). Embora uma infecção por CMV progride assintomaticamente na maioria das pessoas imunocompetentes, pode causar graves complicações em pessoas cujo sistema imunitário esteja debilitado ou ainda não completamente desenvolvido (hepatite, retinite, pneumonia, etc.). O CMV é o agente patogénico mais frequente em infecções congénitas. Aproximadamente 10% das crianças infectadas congenitalmente por CMV apresentam sintomas ao nascimento (icterícia, hepatosplenomegalia, petequia e coriorretinite). Outros grupos de pacientes para quem uma infecção aguda por CMV representa uma grave ameaça são pacientes com transplantes de órgãos e medula óssea, e pacientes com SIDA. O ensaio de antigenemia para CMV foi desenvolvido utilizando um cocktail de dois anticorpos monoclonais (C10/C11) dirigidos contra a pp65 [2]. O ensaio de antigenemia para CMV é importante para o diagnóstico da infecção activa por CMV. As culturas em frasco estéril (shell vial) fornecem resultados após 1 a 2 dias, mas não são sensíveis à detecção de CMV em amostras de sangue. Por oposição, a detecção do antígeno CMV em células (antigenemia CMV) polimorfonucleares de sangue periférico (PMN) é sensível e rápida [1, 2]. Esta técnica utiliza anticorpos monoclonais para detectar a fosfoproteína da matriz inferior do CMV (pp65), um antígeno precoce na replicação de vírus, de presença abundante nos PMNs positivos ao antígeno [2-5]. O ensaio CMV Brite Turbo pode ser concluído em 2 horas após a colheita da amostra.

Princípio do teste - teste de imunofluorescência -

O método CMV Brite™ Turbo consiste em:

- lise directa de eritrócitos de sangue periférico
- preparação de lâminas de Cytospin
- fixação e permeabilização
- coloração de imunofluorescência indireta utilizando anticorpos monoclonais dirigidos contra a proteína pp65 do CMV
- leitura e avaliação dos resultados

O primeiro passo do método CMV Brite™ Turbo implica a lise directa de eritrócitos de sangue periférico [6]. Após a lise, os leucócitos são citocentrífugados numa lâmina, fixados e permeabilizados para permitir a subsequente detecção do antígeno pp65. A presença de pp65 é detectada pelo cocktail de anticorpos C10/C11 e visualizada através de um anticorpo secundário específico rotulado FITC. Os leucócitos positivos ao antígeno CMV apresentam coloração nuclear polilobada amarela-esverdeada homogénea quando observada utilizando um microscópio de fluorescência. O número de células positivas ao antígeno CMV é contabilizado por coloração duplicada.

Kit contém

LYS	200 ml	Reagente A (A) , Solução de lise de eritrócitos (Solução de cloreto de amônio, azida de sódio < 0,1%) Concentrado: Diluir 1:10 com água desmineralizada.
FIX	290 ml	Reagente B (B) , Solução fixadora (formaldeído em PBS, azida de sódio < 0,1%) Concentrado: diluir 1:5 com PBS.
PER	290 ml	Reagente C (C) , Solução de permeabilização (Igepal Ca-630, soro de vitelo recém-nascido em PBS, azida de sódio < 0,1%), Concentrado: diluir 1: 5 com PBS.
MAB	4 ml	Reagente D (D) , Anticorpo monoclonal (rato), Mistura de C10/C11 (IgG/IgG1) contra proteína pp65 de matriz inferior. Azida de sódio < 0,1% (Pronto a utilizar)
CONJ	4 ml	Reagente E (E) , Imunoglobulinas de ovelha anti-rato conjugadas com FITC e Azul de Evans, azida de sódio < 0,1%. (Pronto a utilizar)
CONTR	5 x 1	Lâmina de Controlo , Lâminas de microscópio para controlo de antigenemia CMV. Lâmina de controlo em bolsa selada com disssecante. (Pronto a utilizar)

Material de laboratório necessário mas não incluído no Kit

Centrifugadora de laboratório; tubos de centrifugação com fundo cónico de 50 ml; Pipetas esterilizadas, micropipeta e pontas; solução salina tamponada com fosfato (PBS), pH 7.4, Ca, sem Mg; Hemocitómetro ou contador automático de células; Lâminas de citocentrífugação;

Citocentrífugadora (como o modelo Cytospin 3 de Shandon Southern Products, Ltd.); Jarras de Coplin ou de coloração histológica; Câmara húmida; incubadora a 37 °C; Microscópio de fluorescência com ampliação de 250x a 1000x; Suporte (não fluorescente, como o CITI FLUOR, solução de Glycerol e PBS, UKC Chem Lab; ou Suporte Immunoconcept, Catálogo # 111); Hotte (sistema de filtração); Lâminas; Cronômetro/Temporizador.

Aviso e precauções

Não incorpore reagentes. Evite o contacto com os olhos e a pele. Todas as amostras e materiais utilizados para o teste devem ser considerados potencialmente infecciosos e devem ser tomadas as devidas precauções de segurança. Para a preparação de Lâminas de Controlo do CMV Brite™ Turbo foram utilizados leucócitos obtidos de um dador de sangue humano saudável. Cada amostra de dador foi testada e obteve resultados negativos para a presença de anticorpos para HIV-1, HIV-2, HCV e CMV, bem como para HBsAg. Não pipete com a boca. Em conformidade com as boas práticas laboratoriais, utilize luvas, bata e óculos de proteção. Os líquidos e materiais não combustíveis devem ser descontaminados com hipoclorito de sódio (concentração final: 3%, tempo de actividade de pelo menos 30 minutos). Os resíduos líquidos que contenham ácidos devem ser neutralizados antes de serem eliminados. Todos os materiais reutilizáveis devem ser autoclavados durante 1 hora a 121 °C.

LYS contém cloreto de amônio. H302 Nocivo por ingestão; H319 Provoca irritação ocular grave. P305 + P351 + P338 SE ENTRAR EM CONTACTO COM OS OLHOS: enxaguar cuidadosamente com água durante vários minutos. Se usar lentes de contacto, retire-as, se tal lhe for possível. Continuar a enxaguar. **FIX** contém menos de 9.3% de formaldeído. O formaldeído é um alergénico altamente tóxico e reagente potencialmente carcinogénico, que deve ser manuseado em conformidade com as boas práticas laboratoriais tomando as devidas precauções. Evite o contacto com os olhos e a pele. H301 + H311 + H331 Tóxico por ingestão, contacto com a pele ou inalação; H314 Provoca queimaduras na pele e lesões oculares graves; H317 Pode provocar uma reacção alérgica cutânea; H335 Pode provocar irritação das vias respiratórias; H351 Suspeito de provocar cancro; H370 Afeta os órgãos. P260 Não respirar as poeiras/fumos/gases/névoas/vapores/aerosóis; P280 Usar luvas de protecção/vestuário de protecção/protecção ocular/ protecção facial; P301 + P310 EM CASO DE INGESTÃO: contacte imediatamente um CENTRO DE INFORMAÇÃO ANTIVENENOS ou um médico; P305 + P351 + P338 SE ENTRAR EM CONTACTO COM OS OLHOS: enxaguar cuidadosamente com água durante vários minutos. Se usar lentes de contacto, retire-as, se tal lhe for possível. Continuar a enxaguar; P310 Contacte imediatamente um CENTRO DE INFORMAÇÃO ANTIVENENOS ou um médico. **CONJ** contém Azul de Evans. Evite o contacto com os olhos, a pele e as roupas. Mantenha o recipiente bem fechado. Lave abundantemente com água após contacto com **FIX** e **CONJ** e consulte um médico. O teste deve ser efectuado por técnicos de laboratório qualificados e autorizados. Os testes são efectuados em condições assépticas e de controlo microbiológico. Contacte o fabricante se o kit de testes original estiver danificado.

Armazenamento

Após a recepção, armazene os reagentes entre 2-8 °C. Evite a exposição directa à luz solar. Os reagentes armazenados em conformidade com as instruções de armazenamento indicadas permanecem estáveis até ao fim da data de validade indicada no rótulo. Em caso de testes sucessivos, armazene imediatamente os reagentes após a sua utilização entre 2-8 °C.

Processamento da amostra sanguínea

Proceda à colheita de 3 a 5 ml de sangue venoso num tubo tratado com EDTA, utilizando punção venosa asséptica. Envie a amostra para o laboratório sem demoras. A amostra sanguínea deve ser mantida à temperatura ambiente

(20-25 °C) até ao processamento. O processamento deve ser efectuado entre 6 a 8 horas após a colheita da amostra, dado que foi demonstrado que pode ocorrer uma diminuição das células positivas ao antígeno (colhidas em heparina) após o armazenamento [7-8, 11]. Por isso, deve efectuar-se a preparação imediata das lâminas sempre que possível. Em pacientes com neutropenia grave (quantidade total de neutrófilos inferior a 200/mm³) podem ser necessários pelo menos 10 ml de sangue. Se as amostras forem transportadas, têm de ser acondicionadas de acordo com os requisitos legais para o transporte de materiais infecciosos.

Procedimento de Teste

O protocolo tem de ser seguido à risca. Excepto se mencionado o contrário, os reagentes (incluindo PBS) devem estar à temperatura ambiente quando utilizados no procedimento de teste. A temperatura ambiente está definida entre 20-25 °C. Deixe as lâminas de controle atingirem a temperatura ambiente antes de abri-las.

I. Preparação da suspensão de leucócitos

- Diluir o reagente A 1:10 em água desmineralizada (destilada) e permitir o arrefecimento até 2-8 °C.
- Misturar 2 ml de sangue com 30 ml de solução de lise para eritrócitos diluída fria (2-8 °C) num tubo de fundo cônico de 50 ml e incubar durante 5 minutos a 2-8 °C.
- Centrifugar durante 2 minutos a 1000xg (2500 rpm). Eliminar o sobrenadante.
- Repetir o passo de lise fria se a lise de eritrócitos não for suficiente após a primeira vez.
- Suspender novamente o sedimento de células em 30 ml de PBS.
- Centrifugar durante 2 minutos a 1000xg (2500 rpm). Eliminar o sobrenadante.

- Suspender novamente o sedimento em 1 ml PBS. (Em pacientes com neutropenia grave, pode ser necessário um volume mais baixo, de até 0.2 ml, para a resuspensão).

II. Contagem de células

- Contar células utilizando um Hemocitómetro ou um contador automático de células.
- Ajustar a concentração para 2.0×10^6 células/ml através de diluição em PBS.

III. Preparação das lâminas de citocentrifugação

- Centrifugar 100 µl da suspensão de 2.0×10^6 células/ml a aprox. 600 rpm (54x g) durante 4 minutos em lamelas de vidro por citocentrifugação.
- Preparar, pelo menos, 3 lâminas por amostra de paciente (2 para testar e uma adicional de reserva).
- Deixar secar as lâminas durante aprox. 5 minutos.
- Delimitar a área das células na lâmina utilizando um marcador de laboratório.
- As lâminas devem ser mantidas à temperatura ambiente uma noite antes da fixação.

IV. Fixação e permeabilização

- Diluir a solução fixadora (reagente B) 1: 5 em PBS numa hotte antes da utilização.
- Diluir a solução de permeabilização (reagente C) 1: 5 em PBS numa hotte antes da utilização. (Não reutilizar.)
- Mergulhar 2 lâminas em reagente B diluído durante 5 minutos à temperatura ambiente na hotte.
- Mergulhar as lâminas 3 vezes em PBS (solução de lavagem) e deixar na mesma solução de lavagem durante 3 minutos.
- Mergulhar as lâminas em reagente C diluído durante 1 minuto à temperatura ambiente.
- Mergulhar as lâminas 3 vezes na solução de lavagem e colocar em nova solução de lavagem durante 5 minutos (ou qualquer período até 60 minutos). Se a coloração com anticorpo monoclonal for efectuada em seguida, prossiga para o passo V.

Para armazenar lâminas:

Se armazenar as lâminas, lave-as em água desmineralizada (destilada) durante 15 segundos. Deixar secar as lâminas sob a hotte durante aprox. 20 minutos. Uma vez secas, as lâminas devem ser embaladas em folha de alumínio e armazenadas a 2-8 °C durante 24 horas ou congeladas a -80 °C.

V. Coloração de imunofluorescência

- A partir deste ponto, não permitir a secagem das preparações de células até ao fim do procedimento de coloração.
- Lâmina de controlo CMV Brite™ Turbo. Hidratar novamente a lâmina de controlo em PBS durante 1 ou 2 minutos.
- Remover uma lâmina de cada vez da solução de lavagem e secar cuidadosamente a área em torno da concentração celular.
- Aplicar 35 µl de solução MoAb C10/C11 (reagente D) e incubar durante 20 minutos a 37 °C numa câmara húmida.
- Mergulhar as lâminas 3 vezes em solução de lavagem (PBS) e colocar em solução de lavagem nova durante 3 minutos.
- Remover uma lâmina de cada vez da solução de lavagem e secar cuidadosamente a área em torno da concentração celular.
- Aplicar 35 µl de conjugado (reagente E) e incubar durante 20 minutos a 37 °C numa câmara húmida.
- Lavar duas vezes em PBS novo, lavar cuidadosamente com água da torneira (3 vezes) e colocar em suporte próprio e cobrir com uma lamela.

VI. Leitura

Efectuar a leitura assim que possível. As lâminas podem ser armazenadas até 8 horas a 4 °C se forem cobertas hermeticamente para minimizar a descoloração. Efectuar a avaliação microscópica utilizando um microscópio de imunofluorescência com ampliação a 400x. É possível utilizar uma ampliação superior de 1000x m para aumentar a resolução. (Consultar nota em recomendações para cubo de microscópio e de fluorescência em "Material e equipamento de laboratório"). Analise toda a superfície da concentração. Devem ser observadas duas concentrações por paciente. As células positivas apresentam uma coloração nuclear polilobada amarela-esverdeada homogénea. As células negativas não apresentam uma coloração nuclear amarela-esverdeada. Leituras ambíguas resultantes de artefactos podem ocorrer raramente (menos de 5%).

O artefacto mais comum deve-se a eosinófilos. São reconhecidos por uma coloração citoplasmática não específica, com o núcleo semelhante a buracos negros, frequentemente em forma de óculos. O citoplasma pode apresentar-se parado e mais amarelado. Se todas as células se apresentarem esverdeadas, isso representa um artefacto que pode ser associado a alguns tipos de pacientes, e que é muito raro.

Uma coloração esverdeada de células apenas na periferia da concentração não é considerada como resultado positivo. Isto pode ocorrer se a concentração tiver começado a secar durante a incubação, na solução de anticorpos monoclonais ou na solução conjugada.

Assegurar a incubação numa câmara húmida e verificar se toda a concentração celular está coberta pelo reagente.

Se não for possível interpretar as lâminas devido a leituras ambíguas de artefactos, corar a(blhjanssen fcports) lâmina(s) de reserva; se permanecer a ambiguidade, obter e testar nova amostra.

Controlo de Qualidade

As lâminas de controlo positivo (proteína de matriz inferior de CMV fixa pp65 células positivas para抗ígenos misturados com leucócitos negativos para抗ígenos) e negativo (leucócitos fixos negativos para抗ígenos de CMV) são fornecidas no kit. Os slides são utilizados apenas para verificar o procedimento de coloração e não influenciam o diagnóstico do kit. O controlo positivo deve demonstrar coloração adequada antes da avaliação das amostras dos doentes. O controlo positivo deve apresentar coloração amarelo-verde homogéneo de células positivas com morfologia redonda (núcleo não é visível). O controlo negativo não deve apresentar coloração amarelo-esverdeado. Apesar do facto dos procedimentos de qualidade sejam respeitados rigorosamente, poderiam ser observadas células positivas no controle negativo. Nesses casos, a coloração não deve ser considerada inválida. As células individuais positivas sobre o controle negativo não significa que a lâmina com material de paciente não possa ser interpretada.

Escolha do microscópio

É importante seleccionar a configuração correcta do cubo cujo filtro corresponda ao fluorocromo a utilizar. Isto varia consoante a marca do microscópio. Seleccionar a combinação correcta para utilização de FITC, como o microscópio Olympus BX-40 ou BX-60 com: Cubo de fluorescência (U-MNB), Comprimento de onda de excitação (470-490 nm), Comprimento de onda de emissão (> 515 nm).

Leitura e interpretação do resultado

Os resultados da avaliação das lâminas de amostras de pacientes são qualitativos. Resultado positivo: presença de uma ou mais células positivas ao抗ígeno CMV por coloração duplicada. Resultado negativo: inexistência de células positivas ao抗ígeno CMV por coloração duplicada. Deve existir um mínimo de aprox. 50.000 células na amostra para determinar que um resultado é negativo. Se ocorrem leituras ambíguas devido a artefactos em ambas as colorações duplicadas, os resultados não devem ser interpretados. Corar as lâminas de reserva ou obter e testar nova amostra do paciente.

Limitações do procedimento

A detecção da proteína estrutural precoce pp65, da matriz inferior do CMV, deve ser efectuada por laboratórios com experiência em técnicas imunocitoquímicas.

A preparação de leucócitos deve ser efectuada por pessoal técnico com experiência em técnicas assépticas. A eficácia do CMV Brite™ Turbo Kit com amostras que não leucócitos de sangue humano ainda não foi determinada. As características de desempenho do teste foram determinadas utilizando o CMV Brite™ Kit (amostras de EDTA e heparina) e validadas em estudos internos utilizando o CMV Brite™ Turbo Kit (apenas amostras de EDTA). O CMV Brite™ Turbo Kit destina-se apenas à utilização com microscopia de imunofluorescência e não para utilização com cromatometria de fluxo. O tipo e a condição dos instrumentos utilizados influenciam o aspecto visual da imagem obtida. A reacção pode variar consoante o tipo de microscópio utilizado, fonte de luz, uso da lâmpada, sistema e espessura dos filtros, diferenças na sensibilidade do substrato抗ígenico ou procedimento de ensaio utilizado. Cada laboratório deve estabelecer os próprios critérios de leitura de amostras de pacientes utilizando os controlos adequados. A detecção e confirmação da proteína estrutural precoce pp65, da matriz inferior do CMV, em leucócitos de sangue periférico não constitui diagnóstico de doença sintomática, dado que a pp65 do CMV pode estar presente durante um período de tempo significativo após uma infecção aguda, e pacientes com antigenemia CMV (especialmente baixos níveis) podem ter doença assintomática. Existem muitos estudos publicados de pacientes que têm antigenemia negativa e viremia positiva, e alguns destes pacientes podem ter doença por CMV. Assim sendo, um teste de antigenemia negativa não exclui absolutamente infecção ou doença por CMV. O CMV Brite™ Turbo Kit não se destina à monitorização de medicamentos antivirais. As características de desempenho do teste ainda não foram determinadas utilizando amostras neonatais. Pode ser observada uma diminuição de células positivas ao抗ígeno (colhidas em heparina) por lâmina após o armazenamento [7-8, 11]. Por isso, deve efectuar-se a preparação imediata das lâminas sempre que possível. O tempo máximo de armazenamento de amostras à temperatura ambiente ainda não foi determinado. Em pacientes com neutropenia grave (quantidade total de neutrófilos inferior a 200/mm³) podem ser necessários pelo menos 10 ml de sangue. Dado que os anticorpos monoclonais foram preparados utilizando uma estirpe protótipo, podem não detectar todos os抗ígenos ou novas estirpes de CMV. Por exemplo, os anticorpos monoclonais podem não detectar estirpes de CMV que tenham sofrido pequenas alterações de aminoácidos na região do epitopo alvo. As características de desempenho do teste ainda não foram determinadas utilizando amostras colhidas em heparina.

Características de desempenho

O CMV Brite™ Kit foi comparado à detecção do vírus CMV em cultura convencional (CC) e em cultura em frasco estéril (SV, shell vial) utilizando leucócitos de sangue periférico humano. Foram avaliadas 300 amostras clínicas em comparação com cultura convencional (CC) e cultura em frasco estéril (SV) [9]. Foi colhido um total de 300 amostras de sangue venoso de pacientes transplantados (159 de medula óssea alógena e 47 de órgãos sólidos), 85 pacientes com HIV positivo/SIDA e 9 pacientes imunocompetentes. O CMV Brite™ Kit determinou 92 amostras (30,7%) positivas ao CMV. Em populações de doadores de sangue aleatórios, o número de amostras consideradas repetidamente reactivas à antigenemia CMV pelo CMV Brite™ Kit foi habitualmente 0%. Nota: As características de desempenho foram estabelecidas utilizando o CMV Brite™ Kit e validadas em estudos internos utilizando o CMV Brite™ Kit Turbo Kit.

Comparação do CMV Brite™ Kit com CC e SV

CMV Brite Kit			
CC/SV	+	-	total
+	31	3	34
-	61	205	266
total	92	208	300

Sensibilidade: $31/34 = 91,2\% \text{ (95\% CI} = 76,3 \text{ to } 98,1\%)$ Especificidade: $205/266 = 77,1\% \text{ (95\% CI} = 70,2 \text{ to } 81,1\%)$

O desempenho do CMV Brite™ Turbo Kit foi validado num estudo de 183 amostras de pacientes. As amostras de pacientes incluíam 173 pacientes com transplante de órgãos, 9 pacientes imunocompetentes e 1 amostra de paciente HIV positivo. Cada paciente foi testado em paralelo com o CMV Brite™ Kit e o CMV Brite™ Turbo Kit.

Os resultados deste estudo de validação são fornecidos na tabela seguinte.

CMV Brite™ Kit			
CMV Brite™ Turbo Kit	+	-	Total
+	43	7	50
-	6	127	133
Total	49	134	183

Reactividade cruzada

Foram testados os seguintes isolados clínicos de vírus para a análise de reactividade cruzada: Vírus herpes tipo 1 e tipo 2, Vírus varicela-zoster, Adenovírus 2, 4 e 5, Vírus parainfluenzae 1, 2 e 3, Vírus sincicial respiratório, Poliovírus 3 (tipo selvagem 3), Ecovírus 11, Ecovírus 30, Vírus Epstein-Barr (estirpe laboratorial P3HR1), Vírus herpes humano tipo 6 (estirpe laboratorial GS), Vírus da imunodeficiência humana (tipo 1, lâminas IF disponíveis no mercado). Não foi observada reactividade cruzada, à exceção de uma fraca reacção positiva em seis dos sete isolados HSV-1 no ensaio de frasco estéril. A coloração fraca observada nos seis isolados HSV-1 era citoplásrica (observada apenas fora do núcleo, num número limitado de pequenos focos), similar ao padrão IF habitualmente encontrado com anticorpos monoclonais HSV-1. É possível que a coloração fraca observada seja devida à expressão dos receptores Fc pelas células infectadas com HSV-1 no ensaio de frasco estéril. Uma análise subsequente concluiu não existirem provas de reactividade cruzada entre o HSV e o ensaio de antigenemia para CMV utilizando o cocktail de anticorpos monoclonais C10/C11.

Literatura

- Van der Bij, W., Schirm, J., Torensma, R., Van Son, W.J., Tegzess, A.M., The, T.H. (1988) Comparison between viremia and antigenemia for detection of cytomegalovirus in blood. *J. Clin. Microbiology* **26**, 2531 - 2535.
- Grefte, J.M.M., Van der Gun, B.T.F., Schmolke, S., Van der Giessen, M., Van Son, W.J. and The, T.H. (1992). The lower matrix protein pp65 is the principal viral antigen present in peripheral blood leukocytes during an active cytomegalovirus infection. *J.Gen.Viro.* **73**, 2923-2932.
- Gerna, G., Revello, M.G., Percivalle, E., Zavattoni, M., Pareja, M., Battaglia, M. (1990). Quantification of human cytomegalovirus viremia by using monoclonal antibodies to different viral proteins. *J. Clin. Microbiology* **28**, 2681 - 2688.
- Revello, M.G., Percivalle, E., Di Matteo, A., Morini, F., Gerna, G. (1992). Nuclear expression of the lower matrix protein of human cytomegalovirus in peripheral blood leukocytes of immunocompromised viremic patients. *J. Gen Virol* **73**, 437 - 442.
- Gerna, G., Revello, M.G., Percivalle, E., Morini, F. (1992). Comparison of different immunostaining techniques and monoclonal antibodies to the lower matrix phosphoprotein (pp65) for optimal quantitation of human cytomegalovirus antigenemia. *J. Clin. Microbiology* **30**, 1232 - 1237.
- Ho, S.K.N., Lo, C-Y., Cheng,I.K.P. and Chan,T-M., (1998) Rapid cytomegalovirus pp65 antigenemia assay by direct Erythrocyte Lysis and Immunofluorescence Staining. *J. Clin.Microbiology* **36**, 638-640.
- Boeckh, M., Woogerd, P.M., Stevens-Ayers, T., Ray, C.G., and Bowden, R.A. (1994). Factors influencing detection of quantitative Cytomegalovirus antigenemia. *J.Clin.Microbiology* **32**, 832-834.
- Landry, M.L., Ferguson D., Cohen, S., Huber, K., and Wetherill, P. (1995). Effect of delayed specimen processing on Cytomegalovirus antigenemia test results. *J.Clin.Microbiology* **33**, 257-259.
- Landry, M.L., Ferguson, D., Stevens-Ayers, T., De Jonge, M.W.A., and Boeckh, M. (1996). Evaluation of CMV Brite TM Kit for detection of Cytomegalovirus pp65 antigenemia in peripheral blood leukocytes by immunofluorescence. *J.Clin.Microbiology* **34**, 1337-1339.
- NEN EN ISO 15223-1 dispositivos médicos - Símbolos para ser usado com rótulos de dispositivos médicos, rotulagem e informação a ser fornecida - Parte 1: Requisitos gerais.
- Niubò, J., Pérez, J.L., Carvajal, A., Ardanuy, C. and Martín, R. (1994). Effect of delayed processing of blood samples on performance of cytomegalovirus antigenemia assay. *J.Clin.Microbiology* **32**, 1119-1120.

Estatuto Regulamentar

Neste momento, o CMV Brite™ Turbo Kit é registrado como "in vitro dispositivo médico de diagnóstico" na Brasil, Coréia do Sul, os Estados Unidos da América e nos países que pertencem à Comunidade Européia. Em todos os outros países, devem ser rotulado como "somente para uso em pesquisa".

Explicação dos símbolos utilizados

	Consulte as instruções de utilização
	Referência de catálogo
	Suficiente para
	Dispositivos médicos de diagnóstico in Vitro
	Atenção, consulte a documentação incluída
	Manter afastado da luz (solar)
	Risco biológico
	Limite de temperatura (°C)
	Pesquisa somente para uso
	Código do lote
	Prazo de validade
	Fabricante
	Representante autorizado na Comunidade Europeia
	Conformité Européenne (Conformidade Europeia)

Informações de contato**IQ Products BV**

www.iproducts.nl

Rozenburglaan 13a
9727 DL Groningen
The Netherlands
T +31 (0)50 5757000
F +31 (0)50 5757002
marketing@iproducts.nl

© 2023 - Produtos de IQ Products bv. Todos os direitos reservados. Nenhuma parte dessas obras pode ser reproduzida em qualquer forma, sem permissões.

Current version + release date	Version 1 April 2023
Previous version	N/A
Changes	Created this online version with translations of the international booklet which is printed and included with every kit. The online version includes Greek, Spanish and Italian, whereas the international booklet includes English and Portuguese.
Justification	Environmental reasons.