

PRODUCT INFORMATION SHEET

Monoclonal antibodies detecting human antigens

CD16 FITC **CD19** R-PE **CD3** CyQ RUO REF IQP-423FRC 50 tests

RUO *For Research Use Only*



Description

CD16	Clone	B-E16	Isotype	IgG2a
<i>For detailed description of this particular single reagent, please refer to IQP-130, CD16 (B-E16)</i>				
CD19	Clone	HD37	Isotype	IgG1
<i>For detailed description of this particular single reagent, please refer to IQP-515, CD19 (HD37)</i>				
CD3	Clone	UCHT1	Isotype	IgG1
<i>For detailed description of this particular single reagent, please refer to IQP-519, CD3 (UCHT1)</i>				

Intended use

IQ Products' triple combination IQP-423FRC, CD16 FITC/CD19 R-PE/CD3 CyQ is a 3-color direct immunofluorescence reagent for flow cytometry.

Applications


The triple combination IQP-423FRC, CD16 (NK-cells)/ CD19 (B-cells)/CD3 (T-cells) makes simultaneous and quantitative *in vitro* evaluation of B-, T- and NK-cells possible in peripheral blood.

Usage

All these reagents are effectively formulated for direct immunofluorescent staining of human tissue for flow cytometric analysis using 10 µl/10⁶ leukocytes for singles and 20 µl/10⁶ leukocytes in case of dual and triple combinations. Since applications vary, each investigator should titrate the reagent to obtain optimal results.

Representative Data

Staining with the triple combination CD16 FITC, CD19 R-PE and CD3 CyQ was performed using 20 µl of the conjugated monoclonal antibody and 100 µl of a washed blood sample.



Immunofluorescence staining and lysing protocol

Flow cytometry method for use with dual and triple combinations

1. Add 100 µl of EDTA-treated blood (i.e. approx. 10⁶ leukocytes) to a 5 ml reagent tube. The content of one tube is sufficient to perform one test.

For combinations with anti-kappa and/or anti-lambda Ig see application note below.

2. Add to each tube 20 µl of labeled monoclonal antibody combination*.
3. Vortex the tube to ensure thorough mixing of antibody and cells.
4. Incubate the tube for 15 minutes at room temperature in the dark.
5. Add 100 µl of IQ Lyse (IQP-199 ready-to-use) and mix immediately.
6. Incubate for 10 minutes at room temperature in the dark.
7. Add 2 ml of demineralized water and incubate for 10 minutes in the dark.
8. Centrifuge the labeled cell suspension for 2 minutes at 1000 x g.
9. Remove the supernatant and resuspend the cells in 200 µl of PBS**.
10. Analyze by flow cytometry within four hours (alternatively, the cells may be fixed by 0.05% of formaline in buffered saline for analysis the next day. Some antigens are readily destroyed upon fixation and this should be taken into account when using this alternative).

* Appropriate mouse Ig isotype control samples should always be included in any labeling study

** PBS: Phosphate Buffered Saline, pH 7.2

Application note for anti-kappa and/or anti-lambda Ig combinations

Add 2 ml of PBS containing 0.001% (v/v) Heparin (**prewarmed to 37 °C**) to the cell suspension
Vortex, centrifuge (2 min at 300x g) and discard the supernatant
Repeat this step twice
Resuspend the pelleted blood cells in 100 µl PBS containing 0.001% (v/v) Heparin



Handling and Storage

Antibodies are supplied either as 100 tests per vial (1 ml) for singles or 50 tests per vial (1 ml) for dual and triple combinations. They are supplied in 0.01 M sodium phosphate, 0.15 M NaCl; pH 7.3, 0.2% BSA, 0.09% sodiumazide (NaN₃). Store the vials at 2-8 °C. Monoclonal antibodies should be protected from prolonged exposure to light. Reagents are stable for the period shown on the vial label when stored properly.

Warranty Products sold hereunder are warranted only to conform to the quantity and contents stated on the label at the time of delivery to the customer. There are no warranties, expressed or implied, which extend beyond the description on the label of the product. IQ Products is not liable for property damage, personal injury, or economic loss caused by the product.

Characterization

To ensure consistently high-quality reagents, each batch of monoclonal antibody is tested for conformance with characteristics of a standard reagent. Representative flow cytometric data is included in this data sheet.

Warning All products contain sodiumazide. This chemical is poisonous and hazardous. Handling should be done by trained staff only.

Explanation of used symbols



Consult instructions for use



Catalogue number



Sufficient for



In Vitro Diagnostic medical device



Caution, consult accompanying document



Keep away from (sun)light



Biological risks



Temperature limitation (°C)



For Research Use Only



Batch code



Use by yyyy-mm-dd



Manufacturer



Authorized Representative in the European Community



Conformité Européenne (European Conformity)

		Label - tandem	Ex -max (nm)	Em -max (nm)
P	PURE	purified material	-	-
F	FITC	FITC	488	519
R	R-PE	PE	488, 532	578
C	CyQ	PE-Cy5.18	488, 532	667
A	APC		595, 633, 635, 647	660
PC	PerCP		488, 532	678
PCC	PerCP-Cy5.5		488, 532	695



IQ Products BV
Rozenburglaan 13a
9727 DL Groningen, The Netherlands

+31 (0)50 57 57 000
+31 (0)50 57 57 002
Technical marketing@iqproducts.nl
Orders orders@iqproducts.nl
www.iqproducts.nl